



Shenzhen PTSI Testing Co., Ltd.
http://www.ptsilab.com, E-mail: services@ptsilab.com
Tel: +86 (0) 755-3282 0019 / 4008-516-519, Fax: +86 (0) 755-3221 5519
2/F, Building C, Hongwan Commercial Center, Bao'an Road, Xixiang, Baoan, Shenzhen, China

Certificate of Conformity

NO.: PRSZ22111106E

The following product has been tested by us with the list standards and found in conformity with the council EMC directive 2014/30/EU. It is possible to use CE mark to demonstrate the conformity with this EMC Directive.

Applicant : Shenzhen Micsig Technology Co., Ltd.
Address : 1st Floor, Building A, Huafeng International Robot Industrial Park, Hangcheng Rd, Bao'an, Shenzhen, Guangdong, China, 518126
Manufacturer : Shenzhen Micsig Technology Co., Ltd.
Address : 1st Floor, Building A, Huafeng International Robot Industrial Park, Hangcheng Rd, Bao'an, Shenzhen, Guangdong, China, 518126
Brand : Micsig
Product : Tablet Oscilloscope
Model(s) : TO2004, ATO1002, PTO1002, TO1004, ATO1004, PTO1004, TO2002, ATO2002, PTO2002, TO1002, ATO2004, PTO2004, TO3002, ATO3002, PTO3002, TO3004, ATO3004, PTO3004, TO3502, ATO3502, PTO3502, TO3504, ATO3504, PTO3504, TO5002, ATO5002, PTO5002, TO5004, ATO5004, PTO5004
Standards : EN IEC 61326-1:2021; EN IEC 61326-2-1:2021
EN IEC 61000-3-2:2019+A1:2021;
EN 61000-3-3:2013+A1:2019
Report No. : PQSZ22111106E





.....
(James Liu, Deputy General Manager)
November 25, 2022

This certificate is based on evaluation of one sample of the above-mentioned product. It does not imply an assessment of the whole production.